

**American Industrial Hygiene Association  
Bulk Asbestos Proficiency Analytical Testing Program  
Results of Round A109-416  
1/9/2017**

John Herrock  
University of Louisiana, Monroe - Dept of Toxicology  
700 University Ave.  
Monroe, LA 71209  
UNITED STATES

Laboratory ID Number  
213022

Total Penalty Points 20  
Round Status P  
Program Status P  
Less than 100 penalty points required to pass the round  
Passing 2 of the last 3 rounds required for Program Proficiency

Lot Designation/Sample ID Numbers	A) 1691	B) 3696	C) 6007	D) 3084
<b>Analysis Results from Laboratory</b>				
Asbestos (%)	CHRY (29)	CHRY (2)	NONE	NONE
Other Fibrous Materials (%)	CELL (20)	CELL (23)	CELL (29) FBGL (10)	
Nonfibrous Material (%)	OTHR (43) ACID	ACID (55) OTHR	OTHR (42) ACID	ACID (80) OTHR
Penalty Points Assessed	20	0	0	0
<b>Analysis Results from Reference Laboratory One</b>				
Asbestos (%)	CHRY (65)	CHRY (4)	NONE	NONE
Other Fibrous Materials (%)	CELL (25)	CELL (20)	CELL (30) FBGL (20)	
Nonfibrous Material (%)	OTHR * <sub>1</sub> (10)	PERL (5) OTHR * <sub>3</sub> (5) OTHR * <sub>6</sub> (66)	PERL (35) OTHR * <sub>4</sub> (13) OTHR * <sub>7</sub> (2)	OTHR * <sub>2</sub> (35) OTHR * <sub>5</sub> (65)
<b>Analysis Results from Reference Laboratory Two</b>				
Asbestos (%)	CHRY (55)	CHRY (3)	NONE	NONE
Other Fibrous Materials (%)	CELL (21)	CELL (18)	FBGL (55) CELL (30)	
Nonfibrous Material (%)	OTHR * <sub>8</sub> (9) OTHR * <sub>11</sub> (15)	OTHR * <sub>12</sub> (57) OTHR * <sub>14</sub> (22)	OTHR * <sub>9</sub> (15)	OTHR * <sub>10</sub> (TRA) OTHR * <sub>13</sub> (73) OTHR * <sub>15</sub> (27)
<b>Analysis Results from RTI International</b>				
Asbestos (%)	CHRY (50)	CHRY (2)	NONE	TREM (Tra)
Other Fibrous Materials (%)	CELL (16)	CELL (28)	CELL (28) FBGL (20)	
Nonfibrous Material (%)	OTHR * <sub>16</sub> (34)	OTHR * <sub>17</sub> (32) OTHR * <sub>19</sub> (38)	OTHR * <sub>18</sub> (26) OTHR * <sub>20</sub> (26)	CACO (87) OTHR * <sub>21</sub> (11)